

CERTIFICATE OF CALIBRATION
Standard Reference Source

SRS Number: 102605
Source Description: 5 mL Liquid in Flame Sealed Vial
Product Code: 8125
Customer: LLNS/LLNL for the U.S. DOE/NNSA
P.O. Number: B617058, Item 1

This standard radionuclide source was prepared gravimetrically from a master solution calibrated by Eckert & Ziegler Analytics (EZA). The master solution was calibrated by liquid scintillation counting. Radionuclide calibration and purity were checked by germanium gamma-ray spectrometry, liquid scintillation counting, and/or alpha spectrometry, as applicable. The nuclear decay rate and reference date for this source are given below. EZA maintains traceability to the National Institute of Standards and Technology (NIST) through a Measurements Assurance Program as described in USNRC Regulatory Guide 4.15, Revision 2, July 2007, and compliance with ANSI N42.22-1995, "Traceability of Radioactive Sources to NIST."

Reference Date: 14-March-2016 12:00 PM EST

Isotope	Half-Life, d	Activity, Bq	Uncertainty			Calibration Method**
			$u_A, \%$	$u_B, \%$	$U, \%$ *	
I-125	5.941E+01	1.650E+06	0.2	1.0	2.0	4π LS

Uncertainty:** U - Relative expanded uncertainty, k = 2. See NIST Technical Note 1297, "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results." *Calibration Methods:** 4π LS - 4π Liquid Scintillation Counting, HPGe - High Purity Germanium Gamma-Ray Spectrometer, IC - Ionization Chamber.

(Certificate continued on reverse side)

SRS Number: 102605

Comments:

5.01323 g 0.1 M NaOH + 0.006 M Na₂SO₃ solution with approximately 30 µg/g I carrier

Impurities:

γ-impurities < 0.1%

This source was wipe tested in its inactive areas with leak test results < 185 Bq (5 nCi) of removable activity per ISO 9978:1992.

Source Prepared by: _____



Z. Dimitrova, Radiochemist

QC Approved by: _____



J. Larru, Spectroscopist

Date: 11-MAR-16